



SGS

VERIFICATION OF EN 455 CONDITIONAL COMPLIANCE

No.: SHHL1602007536MD-01C
Product Name: DISPOSABLE NITRILE GLOVE
Style No: XS,S,M,L,XL,XXL

Sufficient samples of the product have been tested and found to be in conformity with
Test Standard: EN455-1:2000 MEDICAL GLOVES FOR SINGLE USE-
PART 1: REQUIREMENTS AND TESTING FOR FREEDOM FROM HOLES
EN455-2:2015 MEDICAL GLOVES FOR SINGLE USE-
PART 2: REQUIREMENTS AND TESTING FOR PHYSICAL PROPRIETIE
EN455-3:2015 MEDICAL GLOVES FOR SINGLE USE-
PART 3: REQUIREMENTS AND TESTING FOR BIOLOGICAL EVALUATION CLAUSE 4.4 & 4.6

as shown in the
Test Report Number(s): SHHL1602007536MD-01

This verification is only valid for the equipment and configuration described, and in conjunction with the test data detailed. It contains the result of the single examination of the subject being in hand and does not represent any universally valid decision concerning the quality of any subject of the current production.

Donna Gu
CRS/Headline SBU Section Head
SGS-CSTC Standards Technical Services Co., Ltd.

Apr 12, 2016

Copyright of this verification is owned by SGS-CSTC Standards Technical Services Co., Ltd. and may not be reproduced other than in full and with the prior approval of the General Manager. This verification is subjected to the governance of the General Conditions of Services, printed overleaf.

Member of SGS Group (Société Générale de Surveillance)

This document is issued, on the Client's behalf, by the Company under its General Conditions of Service printed overleaf. The Client's attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein.

Any other holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not constitute parties to a transaction from exercising all their rights and obligations under the transaction documents.

SGSPAPER
16598818

